

Certificate of Inspection on MITUTOYO Height Master

Date 8. APR. 1987

Code No. 515- 304

Serial No. 006506 Range 0.3 - 18.2"

Tolerance on Length		Accuracy of Lead Screw	
Blocks	Error Millionths	Screw	Error Millionths
1 "	0	.300 "	0
2 "	0	.500 "	-20
3 "	0	.550 "	-15
4 "	+15	.5555 "	-10
5 "	+30	.563 "	-10
6 "	+15	.570 "	-30
7 "	0	.700 "	-25
8 "	+20	.900 "	- 5
9 "	0	1.100 "	+40
10 "	+15	1.300 "	+40
11 "	+10	1.550 "	- - -
12 "	0		
14 "	-30		
16 "	-40		
18 "	-25		
20 "	- - -		
22 "			
24 "			
28 "			
32 "			
36 "			
40 "			

Inspection was performed at 20°C (68°F),
using Standards traceable to the
National Bureau of Standards.
Certificate No.731/237607-86 July 11, 1988

Certified by J. Tamura
Chief Inspector



MITUTOYO MFG. CO., LTD.

31-19, Shiba 5-chome, Minato-Ku, Tokyo, Japan